



Certificated of Inspection

Date: 2010/06/08

Model No.: BM09-DR

Serial No.:

Maximum Capacity: 3-30 N.m

Ambient Temperature: 23°C (73.4°F)

Unit: N.m

1.Torque Value	Reading	Error (%)	2.Torque Value	Reading	Error (%)
3.0	3.0	0.00%	3.0	3.0	0.00%
12.0	12.1	-0.83%	12.0	11.9	0.83%
18.0	17.9	-0.56%	18.0	17.9	0.56%
24.0	24.1	-0.42%	24.0	24.0	0.00%
30.0	30.1	-0.33%	30.0	29.9	0.33%

Quality Manager:

Jack

Inspector:

Derek

Version1. July,2010

Tel: +886-4-8535100

Email: SERVICE@BIRZMAN.COM